

Calibration Report

Laboratory Products Sales

1665 Buffalo Rd./Rochester, N.Y. 14624

Date: 00/00/00

CALIBRATION REPORT

Time: 00:00

Report Type A

Device ID: 7598-400042-VP

Description: 10 μ l

Mode: Addition - Tare

Method ID: LPS METHOD

Description: Standard Single Channel Cal.

Run Date: 00/00/00

Operator:

Balance I.D.: BALANCE #2

STATUS: Passed

Environmental Factors

Temp.: 25.00 °C
 Barom. P.: 760.00 mmHg
 Rel. Hum.: 50.00%

Z Factor: 1.004008
 Air Dens.: 0.001184
 Cub. Exp.: 0.000000

Pipette Tip Information

Part Number: 104
 Tip Type: N/A
 Manufacturer:

Comment:

SUMMARY STATISTICS

Vol.	Ch	Mean	SD	Inaccuracy E%			(not used) Imprecision CV%		
				Actual	Target	Status	Actual	Target	Status
1.0	1	1.0140	0.0116	1.4048	2.50	Passed	1.1433	1.20	N/A
5.0	1	4.9974	0.0050	-0.0510	1.50	Passed	0.1005	0.60	N/A
10.0	1	10.0275	0.0096	0.2753	1.00	Passed	0.0959	0.40	N/A

RUN DETAILS

μ l	Ch	Sp.	Samp. g	Volume
1.0	1	1	0.00102	1.02
1.0	1	2	0.00100	1.00
1.0	1	3	0.00100	1.00
1.0	1	4	0.00102	1.02
5.0	1	1	0.00498	5.00
5.0	1	2	0.00497	4.99
5.0	1	3	0.00498	5.00
5.0	1	4	0.00498	5.00
10.0	1	1	0.00999	10.03
10.0	1	2	0.00998	10.02
10.0	1	3	0.01000	10.04
10.0	1	4	0.00998	10.02